

Product Name: SGA Calibration Standard 3

Part Number: 1404006

Batch Number: C3SGA009


Method of Analysis: Spectrophotometric titration

DataSet Numbers: 06E-1007

pK_a values:

Equation	Measured value	Specification value
$X + H = XH$	9.2	9.2 ± 0.2
$XH + H = XH_2$	5.0	5.0 ± 0.2

UV profile:



This product has been tested by the Quality Control Laboratory of Sirius Analytical Instruments Limited and conforms to the specifications described above.

Approved:

Carolyn Hibbert

.....
Carolyn Hibbert, Quality Assurance Co-ordinator

Certificate Date: 16 May 2006

C3SGA009.doc